

Plot on a graph  $10 \frac{\Delta E}{S}$  ( $y$ -axis) against  $V_S$  ( $x$ -axis) and extrapolate the line obtained until it intersects the  $x$ -axis. At the intersection, the concentration  $C_T$  of the test solution in the ion to be determined is given by the equation:

$$C_T = \frac{C_S V_S}{V_T}$$

#### METHOD III (SINGLE STANDARD ADDITION)

To a volume  $V_T$  of the test solution prepared as prescribed in the monograph, add a volume  $V_S$  of a reference solution containing an amount of the ion to be determined known to give a response situated in the linear part of the calibration curve. Prepare a blank solution in the same conditions. Measure at least three times the potentials of the test solution and the blank solution, before and after adding the reference solution. Calculate the concentration  $C_T$  of the ion to be analysed using the following equation and making the necessary corrections for the blank:

$$C_T = \frac{C_S V_S}{10 \frac{\Delta E}{S} (V_T + V_S) - V_T}$$

- $V_T$  = volume of the test solution or the blank,  
 $C_T$  = concentration of the ion to be determined in the test solution,  
 $V_S$  = added volume of the reference solution,  
 $C_S$  = concentration of the ion to be determined in the reference solution,  
 $\Delta E$  = difference between the average potentials measured before and after adding  $V_S$ ,  
 $S$  = slope of the electrode determined experimentally, at constant temperature, by measuring the difference between the potentials obtained from two reference solutions whose concentrations differ by a factor of ten and are situated within the range where the calibration curve is linear.

01/2008:20237

### 2.2.37. X-RAY FLUORESCENCE SPECTROMETRY<sup>(2)</sup>

Wavelength dispersive X-ray fluorescence spectrometry is a procedure that uses the measurement of the intensity of the fluorescent radiation emitted by an element having an atomic number between 11 and 92 excited by a continuous primary X-ray radiation. The intensity of the fluorescence produced by a given element depends on the concentration of this element in the sample but also on the absorption by the matrix of the incident and fluorescent radiation. At trace levels, where the calibration curve is linear, the intensity of the fluorescent radiation emitted by an element in a given matrix, at a given wavelength, is proportional to the concentration of this element and inversely proportional to the mass absorption coefficient of the matrix at this wavelength.

*Method.* Set and use the instrument in accordance with the instructions given by the manufacturer. Liquid samples are placed directly in the instrument; solid samples are first compressed into pellets, sometimes after mixing with a suitable binder.

To determine the concentration of an element in a sample, it is necessary to measure the net impulse rate produced by one or several standard preparations containing known amounts of this element in given matrices and to calculate or measure the mass absorption coefficient of the matrix of the sample to be analysed.

*Calibration.* From a calibration solution or a series of dilutions of the element to be analysed in various matrices, determine the slope of the calibration curve  $b_0$  from the following equation:

$$b_0 \frac{1}{\mu_M} = \frac{I_C^N}{C}$$

- $\mu_M$  = absorption coefficient of the matrix M, calculated or measured,  
 $I_C^N$  = net impulse rate,  
 $C$  = concentration of the element to be assayed in the standard preparation.

*Mass absorption coefficient of the matrix of the sample.* If the empirical formula of the sample to be analysed is known, calculate its mass absorption coefficient from the known elemental composition and the tabulated elemental mass absorption coefficients. If the elemental composition is unknown, determine the mass absorption coefficient of the sample matrix by measuring the intensity of the scattered X-radiation  $I_U$  (Compton scattering) from the following equation:

$$\frac{1}{\mu_{MP}} = a + b I_U$$

- $\mu_{MP}$  = mass absorption coefficient of the sample,  
 $I_U$  = scattered X-radiation.

*Determination of the net pulse rate of the element to be determined in the sample.* Calculate the net impulse rate  $I_{EP}^N$  of the element to be determined from the measured intensity of the fluorescence line and the intensity of the background line(s), allowing for any tube contaminants present.

*Calculation of the trace content.* If the concentration of the element is in the linear part of the calibration curve, it can be calculated using the following equation:

$$C = \frac{I_{EP}^N}{b_0 \frac{1}{\mu_{MP}}} \times f$$

- $f$  = dilution factor.

01/2008:20238

### 2.2.38. CONDUCTIVITY

The current  $I$  (in amperes) flowing in a conductor is directly proportional to the applied electromotive force  $E$  (in volts) and inversely proportional to the resistance  $R$  (in ohms) of the conductor:

$$I = \frac{E}{R}$$

(2) G. Andermann & M.W. Kemp, Analytical Chemistry 30 1306 (1958). Z.H. Kalman & L. Heller, Analytical Chemistry 34 946 (1962). R.C. Reynolds, Jr., The American Mineralogist 46 1133 (1963). R.O. Müller, Spectrochimica Acta 20 143 (1964). R.O. Müller, Spectrochemische Analyse mit Röntgenfluoreszenz, R. Oldenburg München-Wien (1967).

The conductivity (formerly called specific conductance) of a solution ( $\kappa$ ) is, by definition, the reciprocal of resistivity ( $\rho$ ). Resistivity is defined as the quotient of the electric field and the density of the current. The resistance  $R$  (in  $\Omega$ ) of a conductor of cross-section  $S$  (in  $\text{cm}^2$ ) and length  $L$  (in  $\text{cm}$ ) is given by the expression:

$$R = \rho \frac{L}{S}$$

$$\text{thus: } R = \frac{1}{\kappa} \times \frac{L}{S} \text{ or } \kappa = \frac{1}{R} \times \frac{L}{S}$$

$L/S$  corresponds to the ideal cell constant.

The unit of conductivity in the International System is the siemens per metre ( $\text{S m}^{-1}$ ). In practice, the electrical conductivity of a solution is expressed in siemens per centimetre ( $\text{S cm}^{-1}$ ) or in microsiemens per centimetre ( $\mu\text{S cm}^{-1}$ ). The unit of resistivity in the International System is the ohm-metre ( $\Omega\text{m}$ ). The resistivity of a solution is generally expressed in ohm-centimetres ( $\Omega\text{cm}$ ). Unless otherwise prescribed, the reference temperature for the expression of conductivity or resistivity is  $25^\circ\text{C}$ .

The apparatus and operating procedure described below are applicable to laboratory measurement of conductivity greater than  $10 \mu\text{S cm}^{-1}$ . The measurement of conductivity of water is dealt with in the relevant monographs.

#### APPARATUS

The apparatus used (conductivity meter or resistivity meter) measures the resistance of the column of liquid between the electrodes of the immersed measuring device (conductivity cell). The apparatus is supplied with alternating current to avoid the effects of electrode polarisation. It is equipped with a temperature probe and a temperature compensation device. The conductivity cell contains 2 parallel platinum electrodes coated with platinum black, each with a surface area  $S$ , and separated from the other by a distance  $L$ . Both are generally protected by a glass tube. Other types of cells may also be used.

#### OPERATING PROCEDURE

##### Determination of the cell constant

Choose a conductivity cell that is appropriate for the properties and conductivity of the solution to be examined. The higher the expected conductivity, the higher the cell constant that must be chosen (low  $\rho$ ). Commonly used conductivity cells have cell constants of the order of  $0.1 \text{ cm}^{-1}$ ,  $1 \text{ cm}^{-1}$  and  $10 \text{ cm}^{-1}$ . Use a certified reference material, for example a solution of potassium chloride, that is appropriate for the measurement. The conductivity value of the certified reference material, should be near the expected conductivity value of the solution to be examined. Other certified reference materials may be used especially for cells having a constant of  $0.1 \text{ cm}^{-1}$ . Rinse the cell several times with *distilled water R* and at least twice with the certified reference material used for the determination of the cell constant of the conductivity cell. Measure the resistance of the conductivity cell using the certified reference material at  $25 \pm 1^\circ\text{C}$ . The cell constant  $K_{\text{cell}}$  (in  $\text{cm}^{-1}$ ) depends on the geometry of the conductivity cell and is given by the expression:

$$K_{\text{cell}} = R_{\text{CRM}} \times \kappa_{\text{CRM}}$$

$R_{\text{CRM}}$  = measured resistance, expressed in mega-ohms,

$\kappa_{\text{CRM}}$  = conductivity of the certified reference material solution used, expressed in microsiemens per centimetre.

The measured constant  $K_{\text{cell}}$  of the conductivity cell must be within 5 per cent of the value indicated.

If the determination of the cell constant is carried out at a different temperature than that indicated for the certified reference material, the conductivity value may be calculated from the following expression:

$$\kappa_T = \kappa_{\text{TCRM}} \times [1 + \alpha (T - T_{\text{TCRM}})]$$

$\kappa_T$  = value of conductivity at the different temperature,

$\kappa_{\text{TCRM}}$  = value of conductivity of the certified reference material,

$T$  = temperature set for calibration,

$T_{\text{TCRM}}$  = temperature indicated for the certified reference material,

$\alpha$  = temperature coefficient for the conductivity value of the certified reference material; for potassium chloride  $\alpha = 0.021$ .

##### Determination of the conductivity of the solution to be examined

After calibrating the apparatus with a certified reference material solution, rinse the conductivity cell several times with *distilled water R* and at least twice with the aqueous solution to be examined. Carry out successive measurements as described in the monograph.

01/2008:20239

## 2.2.39. MOLECULAR MASS DISTRIBUTION IN DEXTRANS

Examine by size-exclusion chromatography (2.2.30).

**Test solution.** Dissolve 0.200 g of the substance to be examined in the mobile phase and dilute to 10 ml with the mobile phase.

**Marker solution.** Dissolve 5 mg of *glucose R* and 2 mg of *dextran V<sub>0</sub> CRS* in 1 ml of the mobile phase.

**Calibration solutions.** Dissolve separately in 1 ml of the mobile phase 15 mg of *dextran 4 for calibration CRS*, 15 mg of *dextran 10 for calibration CRS*, 20 mg of *dextran 40 for calibration CRS*, 20 mg of *dextran 70 for calibration CRS* and 20 mg of *dextran 250 for calibration CRS*.

**System suitability solution.** Dissolve either 20 mg of *dextran 40 for performance test CRS* (for dextran 40) or 20 mg of *dextran 60/70 for performance test CRS* (for dextran 60 and dextran 70) in 1 ml of the mobile phase.

The chromatographic procedure may be carried out using:

- a column 0.3 m long and 10 mm in internal diameter, packed with *cross-linked agarose for chromatography R* or a series of columns, 0.3 m long and 10 mm in internal diameter, packed with *polyether hydroxylated gel for chromatography R*,
  - as the mobile phase, at a flow rate of 0.5-1 ml/min, kept constant to  $\pm 1$  per cent per hour, a solution containing 7 g of *anhydrous sodium sulphate R* and 1 g of *chlorobutanol R* in 1 litre of *water R*,
  - as detector a differential refractometer,
  - a 100  $\mu\text{l}$  to 200  $\mu\text{l}$  loop injector,
- maintaining the system at a constant temperature ( $\pm 0.1^\circ\text{C}$ ).

#### CALIBRATION OF THE CHROMATOGRAPHIC SYSTEM

Carry out replicate injections of the chosen volume of the marker solution. The chromatogram shows 2 peaks, the first of which corresponds to *dextran V<sub>0</sub> CRS* and the second of which corresponds to *dextrose R*. From the elution volume